

## ● LCR-900 SPECIFICATIONS:

<b>TEST FREQUENCY</b>		100Hz to 100kHz (5 Range)
<b>BASIC ACCURACY</b>		0.3%
<b>TEST SIGNAL LEVEL</b>		0.6Vrms
<b>DISPLAY RANGE</b>	<b>Impedance R</b>	0.000 $\Omega$ to 9999 M $\Omega$
	<b>Capacitance C</b>	0.000 pF to 9999 F
	<b>Inductance L</b>	0.000 $\mu$ H to 9999 kH
	<b>Equivalent Series Resistance ESR</b>	0.000 $\Omega$ to 9999 $\Omega$
	<b>Equivalent Parallel Resistance Rp</b>	0.000 $\Omega$ to 9999 $\Omega$
	<b>Quality factor Q</b>	0.000 to 9999
	<b>Dissipation factor D</b>	0.000 to 9999
	<b>Phase angle <math>\Theta</math></b>	-90° to 90°
<b>MEASUREMENT PARAMETER</b>		Z, L, C, DCR, ESR, D, Q, $\Theta$
<b>OPERATION SELECT</b>		AUTO/MANUAL
<b>DISPLAY MODE</b>		Value, $\Delta$ , $\Delta\%$ , Sorting, Sorting, Range
<b>INTERFACE(OPTION)</b>		USB
<b>POWER SOURCE</b>		115/230 $\pm$ 10%, 50/60Hz